

Application Serial No: 10/530,725

Responsive to the Office Action mailed on: May 13, 2008

### REMARKS

This Amendment is in response to the Office Action mailed on May 13, 2008.

Claims 1 and 6 are amended and are supported, for example, in the specification on page 2, line 25-page 3, line 21. No new matter is added. Claims 1-13 are pending.

#### §112, Second Paragraph:

Claims 1-13 are rejected as being indefinite. With respect to claim 1, the rejection asserts that the features of "a calculation step for calculating a concentration of a specific component in the sample liquid based on a predetermined relationship between variations of response and variations of wavelength with respect to the reference board whose response varies continuously as the wavelength of light irradiated onto the reference board varies, and the first and second detection results" <sup>are</sup> is not clear. With respect to claim 6, the rejection asserts that the features of "a storage for storing the relationship between variations of response and variations of wavelength with respect to a reference board whose response varies continuously as the wavelength of light irradiated onto the reference board varies" <sup>is</sup> is not clear.

Applicants note that claims 1 and 6 are not directed to light irradiation with multiple wavelengths at the time of actual measurement, but to fluctuations or variations of a wavelength due to temperature fluctuations of a light source. A light source is inevitably subjected to temperature fluctuations, which causes the wavelength of light generated by the same light source <sup>to</sup> also fluctuate within a very small range depending on the particular temperature at a particular time point of measurement. These "variations of wavelength", as recited in claims 1 and 6, can result in an inaccurate calculation of a concentration of a specific component in a sample liquid unless accounted for, such as in claims 1 and 6. For clarification, claims 1 and 6 are each amended to recite that "the variations of the wavelength of irradiated light being caused by environmental temperature changes that also cause fluctuations of response". Withdrawal of this rejection is requested.

Application Serial No: 10/530,725  
Responsive to the Office Action mailed on: May 13, 2008

§102 Rejections:

Claims 1-13 are rejected as being anticipated by Matzinger (US Patent No. 5,780,304). This rejection is traversed.

Claim 1 is directed to an analyzing method that recites, inter alia, a calculation step for calculating a concentration of a specific component in the sample liquid based on a predetermined relationship between variations of response and variations of wavelength with respect to the reference board whose response varies continuously as the wavelength of light irradiated onto the reference board varies, and the first and second detection results, the variations of the wavelength of irradiated light being caused by environmental temperature changes that also cause fluctuations of response.

Matzinger does not disclose or suggest these features. In particular, Matzinger does not disclose, or even contemplate, accounting for variations of wavelength in its calculation step. In contrast, Matzinger discloses that adjustment is made so that each of the different measuring apparatuses provides a proper quantity of light energy (i.e., intensity of light) of each LED that is designed to emit light at a fixed wavelength of 660 nm or 940 nm (see column 11, lines 16-18 and column 12, lines 25-56). Matzinger also includes a gray target (45) and a standard zone (60), in which the gray target (45) is not dependent on wavelength. Also, the standard zone (60) only provides a higher reflectance than a reaction zone, and wavelength dependence of the standard zone (60) is not used for measurement correction (see column 9, line 66-page 10, line 21). Thus, any adjustment or correction made to the method is based on the assumption that each LED emits light at a fixed wavelength.

Accordingly, nowhere does Matzinger disclose or suggest variations of wavelength. Thus, Matzinger also cannot disclose or suggest a calculation step for calculating a concentration of a specific component in the sample liquid based on a predetermined relationship between variations of response and variations of wavelength with respect to a reference board whose response varies continuously as the wavelength of light irradiated onto the reference board varies, and the first and second detection results, variations of the wavelength of irradiated light being caused by environmental temperature changes that also cause fluctuations of response. An advantage of these features is that an accurate optical measurement of the concentration of a specific

Application Serial No: 10/530,725

Responsive to the Office Action mailed on: May 13, 2008

component in a sample liquid can be obtained even if the wavelength of the light irradiated onto a reaction system deviates from an intended value. For at least these reasons claim 1 is not suggested by Matzinger. Claims 2-5 depend from claim 1 and should be allowed for at least the same reasons.

Claim 6 is directed to an analyzing device that requires, among other features, a storage for storing the relationship between variations of response and variations of wavelength with respect to a reference board whose response varies continuously as the wavelength of light irradiated onto the reference board varies, variations of the wavelength of irradiated light being caused by environmental temperature changes that also cause fluctuations of response.

Matzinger does not disclose or suggest these features. In particular, Matzinger does not disclose, or even contemplate, variations of wavelength. As discussed above, in Matzinger, adjustment is made so that each of different measuring apparatuses provides a proper quantity of light energy (i.e., intensity of light) of each LED which is designed to emit light at a fixed wavelength of 660 nm or 940 nm (see column 11, lines 16-18 and column 12, lines 25-56). Matzinger also includes a gray target (45) and a standard zone (60). However, the gray target (45) is not dependent on wavelength. The standard zone (60) only provides a higher reflectance than a reaction zone, and wavelength dependence of the standard zone (60) is not used for measurement correction (see column 9, line 66-page 10, line 21). Thus, any adjustment or correction made to the method is based on the assumption that each LED emits light at a fixed wavelength.

Accordingly, Matzinger does not disclose or suggest variations of wavelength. Thus, Matzinger also does not disclose or suggest a storage for storing the relationship between variations of response and variations of wavelength with respect to a reference board whose response varies continuously as the wavelength of light irradiated onto the reference board varies, variations of the wavelength of irradiated light being caused by environmental temperature changes that also cause fluctuations of response. For at least these reasons claim 6 is not suggested by Matzinger. Claims 7-13 depend from claim 6 and should be allowed for at least the same reasons.

Application Serial No: 10/530,725

Responsive to the Office Action mailed on: May 13, 2008

Conclusion:

Applicant respectfully asserts that claims 1-13 are in condition for allowance. If a telephone conference would be helpful in resolving any issues concerning this communication, please contact Applicant's primary attorney-of record, Douglas P. Mueller (Reg. No. 30,300), at (612) 455-3804.



Dated: June 22, 2009

Respectfully submitted,

HAMRE, SCHUMANN, MUELLER &  
LARSON, P.C.  
P.O. Box 2902-0902  
Minneapolis, MN 55402-0902  
(612) 455-3800

By: 

Douglas P. Mueller  
Reg. No. 30,300  
DPM/ahk